



Human vWF-A2 ELISA Kit

Catalog number: NR-E11136 (96 wells)

The kit is designed to quantitatively detect the levels of human von Willebrand Factor A2 Domain (vWF-A2) in serum, cell culture supernatant and other suitable sample solutions.

FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC PURPOSES

Important notes

Before using this product, please read this manual carefully; after reading the subsequent contents of this manual, please note the following specially:

- The operation should be carried out in strict accordance with the provided instructions.
- Store the unused strips in a sealed foil bag at 2-8°C.
- Always avoid foaming when mixing or reconstituting protein solutions.
- Pipette reagents and samples into the center of each well, avoid bubbles.
- The samples should be transferred into the assay wells within 15 minutes of dilution.
- We recommend that all standards, testing samples are tested in duplicate.
- Using serial diluted sample is recommended for first test to get the best dilution factor.
- If the blue color develops too light after 15 minutes incubation with the substrate, it may be appropriate to extend the incubation time (Do not over-develop).
- Avoid cross-contamination by changing tips, using separate reservoirs for each reagent.
- Avoid using the suction head without extensive wash.
- Do not mix the reagents from different batches.
- Stop Solution should be added in the same order of the Substrate Solution.
- TMB developing agent is light-sensitive. Avoid prolonged exposure to the light.


Intended use

The kit is used to quantify the human vWF-A2 in cell culture supernatant, serum, and other suitable sample solution. Does not show cross-reactivity with human pro-Collagen I α 1, pro-Collagen III α 1, glycoprotein V, integrin α V β 3, thrombospondin-1, ADAMTS13 and CD42b

Standard range	46.9 – 3000 pg/ml
Assay time	4 h 40 min
Validity	Six months
Store at	2-8 °C

Assay principle

The human vWF-A2 ELISA Kit is based on standard sandwich enzyme-linked immunosorbent assay technology. Rat anti-human vWF-A2 specific antibody has been pre-coated onto 96-well plate. Human vWF-A2 present in the standards/ samples bind to the capture antibody. Subsequently, biotinylated goat anti-human vWF-A2 detection antibody is added to form an Ab-Ag-Ab sandwich. Streptavidin-HRP is added and unbound conjugate is removed with wash Buffer. Next, addition of HRP substrate, TMB, results in the production of a blue colored product that changes to yellow after the addition of acidic Stop Solution. The density of yellow color is directly proportional to the amount of human vWF-A2 captured on plate.



Materials supplied

1. Human vWF-A2 standard:	3 ng/vial ×2
2. 96-well plate pre-coated with anti-human vWF-A2 Ab:	1
3. Sample Diluent buffer :	12 ml X 2
4. Detection antibody:	1 vial, dilution 1:180
5. Streptavidin-HRP:	1 vial dilution 1:200
6. Antibody Diluent Buffer	12 ml
7. Streptavidin-HRP Diluent Buffer	12 ml
8. Chromogen Solution A:	6 ml
9. Chromogen Solution B:	6 ml
10. Stop Solution:	6 ml
11. 20 × Wash Buffer:	25 ml
12. Plate sealer	2
13. Package insert	1

Materials required but not supplied

- Deionized water.
- Standard plate reader capable of measuring absorbance at 450 nm.
- Adjustable pipettes and disposable pipette tips.
- Multi-channel pipettes, manifold dispenser or automated microplate washer.
- Distilled water.
- Absorbent paper.
- Materials used for sample preparation.

Sample Preparation and storage

Store samples to be assayed within 24 hours at 2-8°C. For long-term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.

- Cell culture supernatant, tissue lysate or body fluids: Remove particulates by centrifugation, analyze immediately or aliquot and store at -20°C
- Serum: Allow the serum to clot in a serum separator tube (about 4hours) at room temperature. Centrifuge at approximately 1000 X g for 15 min. Analyze the serum immediately or aliquot and store frozen at -20°C.
- Plasma: Collect plasma using heparin as an anticoagulant. Centrifuge for 15 min at 1000 x g within 30 minutes of collection. Analyze immediately or aliquot and store frozen at -20°C. EDTA and citrate are not recommended as the anticoagulant.

Reagent Preparation

Standard

- Human vWF-A2: Standard solution should be prepared no more than 2 hours prior to the experiment. Two tubes of standard (3 ng /vial) are included in each kit. Use one tube for each experiment.
- Prepare 3000 pg/ml→46.9 pg/ml of human vWF-A2 standard solutions:
- Add 1 ml of Sample Diluent Buffer into one standard vial with 3 ng human vWF-A2. Keep the tube at room temperature for 10 minutes and mix thoroughly. This is 3000 pg/ml standard solution.
- Label 6 Eppendorf tubes with 1500 pg/ml, 750 pg/ml, 375 pg/ml, 187.5 pg/ml, 93.75 pg/ml, 46.9 pg/ml, respectively. Then make 2-fold serial dilution from 3000 pg/ml to 46.9 pg/ml in 1.5 ml tubes with sample diluent buffer.
- Make sure each tube has $\geq 250 \mu\text{l}$ of standard.

Note: The standard solutions are best used within 2 hours.

Preparation of detection anti-human vWF-A2 antibody working solution

- The stock solution is stable at 2-8 °C for up to 6 weeks. For long-term storage of up to 6 months, aliquot and store at -20 to -70 °C. Avoid freeze-thaw.
- The working solution should be prepared no more than 2 hours prior to the experiment
- The reagent is supplied as 180X concentrate. Empty the total contents in to 11.64 ml of Antibody Diluent Buffer or prepare the solution separately in a volume as needed. The solution should be mixed thoroughly.
- The total volume should be: 0.1 ml/well x the number of wells (Allowing 0.1-0.2 ml more than total volume).

Preparation of Streptavidin-HRP working solution

- The solution should be prepared no more than 1 hour prior to the experiment.
- The total volume should be: 0.1 ml/well x the number of wells (allowing 0.1-0.2 ml more than total volume).
- Streptavidin-HRP should be diluted 1:200 with Streptavidin-HRP Diluent buffer and mixed thoroughly.

Preparation of TMB developing agent working solution

Mix equal volumes of Chromogen Solution A and Chromogen Solution B. The total volume should be: 0.1 ml/well x the number of wells (allowing 0.1-0.2 ml more than total volume). Use it immediately (light sensitive reagent).

Wash Buffer

- If crystals have formed in the 20X wash buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
- Dilute 25 ml Wash Buffer Concentrate (20X) to a total volume of 500 ml with distilled water.

Assay Procedure

Bring all reagents to room temperature before use. Human vWF-A2 standard curve should be prepared for each experiment. The user will decide sample dilution factor by rough estimation of human vWF-A2 concentration in samples.

1. Add 100 μ l of sample or standards per well. Add 0.1 ml of the sample diluent into the control well (Zero well). Cover with an adhesive strip and incubate 2 hours at room temperature.
Note: We recommend that each human vWF-A2 standard solution and each sample is measured in duplicate.
2. Aspirate each well and wash with Wash Buffer, repeating the process two times for a total of three washes. Wash by filling each well with Wash Buffer (300 μ l) using a squirt bottle, manifold dispenser, or auto-washer. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining Wash Buffer by aspirating or by inverting the plate and blotting it against clean paper towels.
3. Add 100 μ l of the Detection Antibody working solution to each well. Cover with a new adhesive strip and incubate for 2 hours at room temperature.
4. Repeat the aspiration/wash as in step 2.
5. Add 100 μ l of the working solution of Streptavidin-HRP to each well. Cover the plate and incubate for 30 minutes at room temperature. Avoid placing the plate in direct light.
6. Repeat the aspiration/wash as in step 2 for three times.
7. Add 100 μ l of TMB developing agent to each well. Cover and incubate at room temperature until a gradient develops and you see visible color in the lowest concentration well (approximately 10-25 minutes depending on the room temperature). Protect from light. Do not over-develop.
8. Add 50 μ l Stop Solution to each well. Mix well.
9. Read the Optical Density (O.D.) at 450 nm using a microtiter plate reader immediately.


Result calculation

For calculation, (the relative O.D.450) = (the O.D.450 of each well) – (the O.D.450 of Zero well). The standard curve can be plotted as the relative O.D.450 of each standard solution (Y) vs. the respective concentration of the standard solution (X). The human vWF-A2 concentration of the samples can be interpolated from the standard curve.

Note: if the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution

Typical data:

This standard curve was generated at the Novatein Biosciences laboratory for demonstration purpose only. A standard curve must be run with each assay.



Manufactured and Distributed by:

Novatein Biosciences

310 W Cummings Park

Woburn, MA, 01801, USA

Phone: (617) 238-1396

Fax: (617) 380-0053

Toll Free: (888) 856-2858

<http://www.novateinbio.com/>

Email: Info@novateinbio.com